

CY4500-EPR EZ-PD™ Protocol Analyzer Utility user guide

About this document

Scope and purpose

This user guide describes the features of EZ-PD™ Protocol Analyzer Utility. The document also provides instructions on the installation and usage of EZ-PD™ Protocol Analyzer Utility.

Intended audience

This document is intended for anyone using the CY4500-EPR kit.

Table of contents**Table of contents**

About this document	1
Table of contents	2
1 The board at a glance.....	3
1.1 Scope of supply	3
1.2 Block diagram.....	3
1.3 Main features	4
1.4 Board parameters and technical data.....	4
2 Installation	6
2.1 Windows installation.....	6
3 Analyzer operation	12
3.1 Capturing PD packets using CY4500-EPR EZ-PD™ Protocol Analyzer	12
3.2 Updating PSOC™ 5LP device firmware on the CY4500-EPR EZ-PD™ Protocol Analyzer	15
3.2.1 Updating firmware using EZ-PD™ Analyzer Utility for Windows	15
References	18
Glossary	19
Revision history	21
Disclaimer	22

The board at a glance

1 The board at a glance

The CY4500-EPR EZ-PD™ Protocol Analyzer supports USB-PD protocol analysis and USB Type-C specifications. It supports the following power ranges:

- Standard power range (SPR) – 21 V at 5 A
- Extended power range (EPR) – 48 V at 5 A

It performs nonintrusive probing and captures accurate protocol messages on the CC line. This analyzer consists of Infineon's programmable MCU (PSOC™ 5LP), which monitors data on the CC lines and sends this data to the host application over a USB interface. The Type-C plug and Type-C receptacle on this analyzer provide a pass-through for the Power Delivery (PD) packets transmitted between each Type-C PD connection. The MCU processor taps these PD packets without disturbing the system and transfers them over the USB interface to a PC running the host application.

Note: EZ-PD™ Analyzer Utility supports decoding of PD packets per the [USB PD Specification Revision 3.2, V1.0](#).

1.1 Scope of supply

The CY4500-EPR EZ-PD™ Protocol Analyzer consists of the following contents:

- CY4500-EPR EZ-PD™ Protocol Analyzer board enclosed in a transparent casing.
- USB 2.0 Type-A to Micro-B Cable
- Six jumper wires
- Quick Start Guide

1.2 Block diagram

The following image shows the block diagram of the CY4500-EPR EZ-PD™ Protocol Analyzer. The PSOC™ 5LP controller decodes the USB-PD messages on the CC lines in a nonintrusive manner, and to measure the voltage and current flowing through the USB interface. These messages are displayed on the EZ PD™ Analyzer Utility application.




Figure 1 Block diagram

The board at a glance

1.3 Main features

- Supports power measurement in the EPR range (48 V at 5 A)
- Captures protocol message on CC lines
- Capable of decoding both the SPR/EPR messages

1.4 Board parameters and technical data

The board comprises a Type-C plug at one end and a Type-C receptacle at the other, which connects to the Type-C device or system under test. The board is powered over the USB micro-B port, which provides 5 V, 500 mA. The board also consists of one LED to indicate the power status. Besides this, it has connector headers brought out for the GPIOs, CC lines, VBUS, and SBU lines from the USB-PD system.

Figure 2, Figure 3, and Figure 4 show the front view, back view, and side view respectively of the CY4500-EPR EZ-PD™ Protocol Analyzer board. Table 1 contains their detailed description. Headers J1 and J8 are located on the back side of the board. Ensure that the exposed header pins of J1 and J8 do not get shorted. Shorting these exposed pins may result in a functional failure or unexpected behavior of the CY4500-EPR EZ-PD™ Protocol Analyzer.




Figure 2 CY4500-EPR EZ-PD™ Protocol Analyzer board (front view)



Figure 3 CY4500-EPR EZ-PD™ Protocol Analyzer board (rear view)

The board at a glance




Figure 4 CY4500-EPR EZ-PD™ Protocol Analyzer board (side view)

The following table contains the details about the pinouts and the connector details of the connectors present on CY4500-EPR board. Connectors J1 and J8 can be used to probe the Type-C signals and the message trigger signals using the provided jumper wires.

Table 1 Connector details

Connector	Description
J1	USB Type-C signal header Pin 1: CC1 Pin 2: CC2 Pin 3: SBU1 Pin 4: SBU2 Pin 5: VBUS Pin 6: GND
J2	Type-C receptacle
J3	Type-C plug
J7	USB Micro-B connector
J8	Message trigger header Pin 1: SOM Pin 2: EOM Pin 3: MTR

- Note:**
- Header J8 pins are used for triggering. The trigger functionality requires EZ-PD™ Protocol Analyzer Utility support. Contact Infineon Sales for more details.
 - The EZ-PD™ Protocol Analyzer board is protected using a transparent casing. Do not remove the board from the casing while using the Analyzer.
 - If multiple CY4500-EPR devices are connected to the laptop, press the tick icon on the EZ-PD™ Protocol Analyzer Utility. This makes the LED on the connected device blink in red color.

Installation

2 Installation

2.1 Windows installation

To install the EZ-PD™ Protocol Analyzer software in Windows, do the following:

1. Download the latest EZ-PD™ Protocol Analyzer Utility software setup file from [EZ-PD™ Protocol Analyzer Utility](#). This package contains the software for running the analyzer and the relevant documentation (user guides, Quick Start Guides, and release notes). Double-click on the executable to start the installation. Read the license agreement and click the “I accept the agreement” radio button when the screen shown in [Figure 5](#) appears.




Figure 5 Install the software – Accept the license agreement

Installation

2. Select the required installation location and click **Next** to start the installation. The default location is *C:\Infineon\Tools\EZ-PD Protocol Analyzer Utility*.




Figure 6 Install the software – Select the installation folder

Installation

3. Check the “Create a desktop” shortcut if you want to create a shortcut on the desktop and click **Next**.




Figure 7 Install the software – Create a desktop shortcut

Installation

4. [Figure 8](#) shows the installation progress.




Figure 8 Install the software – Installation in progress

Installation

5. Check the following buttons if you want to launch the application or view the user guide and click **Finish**.




Figure 9 Install the software – Complete the installation

Installation

6. After the installation is complete, the contents are available at the following location:
<Install location>\EZ-PD Protocol Analyzer Utility.




Figure 10 EZ-PD™ Protocol Analyzer Utility window

Note: For EZ-PD™ Protocol Analyzer Utility support on Mac contact Infineon Sales for more details.

Analyzer operation

3 Analyzer operation

This section describes how to run the CY4500-EPR EZ-PD™ Protocol Analyzer Utility, and how to update the firmware of the PSoC™ 5LP device of the analyzer. Complete the installation as explained in the [Installation](#) section before proceeding.

3.1 Capturing PD packets using CY4500-EPR EZ-PD™ Protocol Analyzer

The steps to run the CY4500-EPR EZ-PD™ Protocol Analyzer to capture the PD packets on the CC bus are as follows:

1. Using a USB Micro-B cable, connect the USB Micro-B receptacle (Connector J7) of the CY4500-EPR EZ-PD™ Protocol Analyzer board to the host PC. LED1 of the CY4500-EPR EZ-PD™ Protocol Analyzer board blinks in white color.
2. Connect the USB Type-C plug (Connector J3) of the CY4500-EPR EZ-PD™ Protocol Analyzer board to the Type-C host device. In this user guide, the [CY7113 - EZ-PD™ PMG1-S3 prototyping kit](#) (not provided with this analyzer) is used as a Type-C PD sink device as an example.
3. Verify that the setup looks similar to the image shown in [Figure 11](#).




Figure 11 Setup for operating CY4500-EPR EZ-PD™ Protocol Analyzer

Analyzer operation

- On the host PC, launch the EZ-PD™ Analyzer Utility from **Start > All Programs > Infineon EZ-PD™ Analyzer Utility**. The EZ-PD™ Analyzer Utility's start-up page looks similar to the image shown in [Figure 12](#).




Figure 12 EZ-PD™ Analyzer Utility start-up page

- Click on the **Start Capturing** icon  to start the CC packet capturing (USB-PD source and sink must be connected after the capture has started or some messages can go missing if we capture after the connection is done).

The Start capturing icon then changes to the **Stop Capturing** icon  once the capturing starts. Alternatively, you can also start capturing data by clicking on **Actions > Start Capturing** from the menu bar. The status bar at the bottom of the utility window shows the message “started capturing” (similar to the status bar at the bottom of [Figure 12](#)).

- Once the capture starts, LED1 on the CY4500-EPR EZ-PD™ Protocol Analyzer will turn green to indicate the same.
- Connect a USB Type-C power adapter (not provided with the analyzer) to the USB Type-C receptacle (Connector J2) of the CY4500-EPR EZ-PD™ Protocol Analyzer board. Verify that the setup looks similar to the image shown in [Figure 13](#). The USB Type-C power adapter mentioned here is just an example. Any USB Type-C device can be used in its place.

Analyzer operation




Figure 13 Complete setup of CY4500-EPR EZ-PD™ Protocol Analyzer

8. Upon connecting the USB Type-C power adapter, the utility running on the host PC automatically displays a string of messages on the screen. Make sure that the hardware connections are not loose or do not get disconnected in the middle of a data capturing event. The EZ-PD™ Analyzer Utility will display messages as shown in Figure 14.




Figure 14 EZ-PD™ Analyzer Utility showing PD packets captured on the CC line

Analyzer operation

9. The EZ-PD™ Analyzer Utility displays the captured messages (sent and received) between the USB Type-C host device and the USB Type-C client device on the CC line. Click on the  icon to stop capturing messages. Alternatively, you can stop capturing data by clicking on **Actions > Stop Capturing** from the menu bar. Upon clicking this icon, the status bar at the bottom of the utility window displays **EZ-PD™ Analyzer is stopped**.
10. Save the captured set of messages by clicking on **File > Save**. These files can be saved in the .ccgx, .xlsx, or .csv formats, but only the file saved in the .ccgx format can be opened using the EZ-PD™ Analyzer Utility by clicking **File > Open**.
11. To clear a captured group of messages, click either the **Clear Data**  icon or click **Actions > Clear Data** from the menu bar. This clears all the captured group of messages. If they are not saved prior to clicking the Clear Data icon, then these messages are lost.
12. If the hardware shown in [Figure 13](#) is not stable or gets disconnected at any point, the status bar at the bottom of the utility window shows the message **EZ-PD™ Analyzer is disconnected**.
13. For more information on how to use the EZ-PD™ Analyzer Utility, see the EZ-PD™ Analyzer Utility User Manual by clicking **Help > User Manual**.

3.2 Updating PSOC™ 5LP device firmware on the CY4500-EPR EZ-PD™ Protocol Analyzer

The PSOC™ 5LP device firmware, present on the CY4500-EPR EZ-PD™ Protocol Analyzer board, can be updated using the **Firmware_Update_Tool Utility**.

3.2.1 Updating firmware using EZ-PD™ Analyzer Utility for Windows

1. Using a USB Micro-B cable, connect the USB Micro-B receptacle (Connector J7) of the CY4500-EPR EZ-PD™ Protocol Analyzer board to the host PC. LED1 of the CY4500-EPR EZ-PD™ Protocol Analyzer board blinks in white color.
2. On the host PC, launch the download firmware application from <install location>\EZ-PD Protocol Analyzer Utility\Firmware_Update_Tool\EZPD_Firmware_Update_Tool. The utility's start-up page looks similar to the image shown in [Figure 15](#).
3. Check the current firmware loaded on the EZ-PD™ Protocol Analyzer hardware by selecting **Get Firmware Version** as shown in [Figure 15](#). A dialog box showing the current firmware version appears.




Figure 15 Get the current firmware version and dialog box

Analyzer operation

- Click **OK** to return to the main menu. Proceed with the rest of the procedure to download the firmware if a new version of the firmware is available.
- Upon executing step 4, the **Download Firmware** window opens as shown in Figure 16. The device enters the bootloader mode by default.




Figure 16 “Download Firmware” window

- Click **Load File...** as shown in Figure 16 and select the CY4500-EPR firmware file in .cyacd format as shown in Figure 17. Click **Open**. The binary file provided with the analyzer is available at the following location:

CY4500-EPR device firmware:

<Install location>\EZ-PD Protocol Analyzer Utility\Firmware_Update_Tool\CY4500_EPR_Firmware

CY4500 device firmware:

<Install Directory>\EZ-PD Protocol Analyzer Utility\Firmware_Update_Tool\CY4500_Firmware




Figure 17 Selecting the CY4500-EPR firmware file

Analyzer operation

- Click **Program** to download the firmware to the device.




Figure 18 Programming the CY4500-EPR firmware file

- Once the message comes up that the download ended, reset the device by disconnecting the Micro-B cable and connecting again.

References

References

- [1] [USB PD Specification Revision 3.2, V1.0](#)

Glossary

Glossary

CC

configuration channel

EoM

end of message

EPR

Extended Power Range

GPIO

general-purpose input output

LED

light emitting diode

Msg. ID

message identification

MTR

message trigger

MCU

microcontroller unit

Obj. Count

object count

PD

Power Delivery

QSG

Quick Start Guide

SPR

Standard Power Range

Glossary

SOM

start of message

SOP

start of packet

SBU

sideband use



Revision history

Revision history

Document revision	Date	Description of changes
**	2024-08-29	Initial release

Trademarks

All referenced product or service names and trademarks are the property of their respective owners.

Edition 2024-08-29

Published by

Infineon Technologies AG
81726 Munich, Germany

© 2024 Infineon Technologies AG.
All Rights Reserved.

Do you have a question about this document?

Email: erratum@infineon.com

Document reference

002-39580 Rev. **

Important notice

The information given in this document shall in no event be regarded as a guarantee of conditions or characteristics ("Beschaffenheitsgarantie")

With respect to any examples, hints or any typical values stated herein and/or any information regarding the application of the product, Infineon Technologies hereby disclaims any and all warranties and liabilities of any kind, including without limitation warranties of non-infringement of intellectual property rights of any third party.

In addition, any information given in this document is subject to customer's compliance with its obligations stated in this document and any applicable legal requirements, norms and standards concerning customer's products and any use of the product of Infineon Technologies in customer's applications.

The data contained in this document is exclusively intended for technically trained staff. It is the responsibility of customer's technical departments to evaluate the suitability of the product for the intended application and the completeness of the product information given in this document with respect to such application.

Warnings

Due to technical requirements products may contain dangerous substances. For information on the types in question please contact your nearest Infineon Technologies office.

Except as otherwise explicitly approved by Infineon Technologies in a written document signed by authorized representatives of Infineon Technologies, Infineon Technologies' products may not be used in any applications where a failure of the product or any consequences of the use thereof can reasonably be expected to result in personal injury.